

**Business Stream Products
Certification Department**

TÜV Rheinland LGA Products GmbH · 90431 Nürnberg

MEAN WELL Enterprises
Co., Ltd.
No. 28, Wuquan 3rd Rd.
Wugu District
NEW TAIPEI CITY 24891
TAIWAN

Contact

Tel. +49 911 655-5225
Mail service@de.tuv.com

Date July 04, 2019

Application for : **CB-Zertifikat**
Certificate No. : DE 02025422
Device : External Switching PSU
Switching Adaptor
Type : see Certificate
Test requirement : IEC 60601-1:2005+A1
IEC 60601-1-6:2010+A1
IEC 60601-1-11:2015
for national deviations see test report

Dear Madame or Sir,

The submitted sample of the product has been tested and in this configuration found to be in accordance with the above mentioned requirements.
Enclosed please find the certificate No. DE 02025422.

Kind regards

Certification body



Dipl.-Ing. (FH) A. Klinker

Test sample: no, documentation available

TÜV Rheinland
LGA Products GmbH

Tillystraße 2
90431 Nürnberg

Tel. +49 911 655-5225
Fax +49 911 655-5226
Mail service@de.tuv.com
Web www.tuv.com/safety

Board of Management

Dipl.-Ing.
Jörg Mähler, Spokesman

Dipl.-Kfm.
Dr. Jörg Schlösser

Chairman of the
Supervisory Board

Dipl.-Ing.
Ralf Scheller

Nuremberg HRB 26013
VAT No.: DE 811835490

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Adaptor
Name and address of the applicant	MEAN WELL Enterprises Co.,Ltd. No. 28, Wuquan 3rd Rd. Wugu District, New Taipei City 24891 Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co.,Ltd. No. 28, Wuquan 3rd Rd. Wugu District, New Taipei City 24891 Taiwan
Name and address of the factory	(See appendix for factories information)
Ratings and principal characteristics	AC 100-240V; 50/60Hz; 0.5-0.25A for model GSM18YZ; 0.7-0.35A for model GSM25YZ; 0.9-0.45A for model GSM36YZ; Class II; for output ratings refer to test report no. 50220419 001 page 9-10
Trademark (if any)	MEAN WELL
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	GSMXYZ (X = 18, 25 or 36 to denote different output power; Y = E, B or U to denote different construction of supply connection terminal; Z = 05, 07, 09, 12, 15, 18, 24, 48 to denote different output voltage)
Additional information (if necessary may also be reported on page 2)	Models differ in model name, rating (input and output), plug construction, windings (different gauge, turns for different output ratings), the ratings of components, HS1 and HS2; for details see test report
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 IEC 60601-1-6:2010+A1 IEC 60601-1-11:2015 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	50220419 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 04.07.2019

Signature:


Dipl.-Ing. (FH) A. Klinker

1. MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.
Wugu District
New Taipei City 24891, Taiwan
2. Suzhou MEAN WELL Technology Co., Ltd. No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park Huang-dai Town, Xiang-cheng District Suzhou, 215152 Jiangsu, China
3. MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD HUADU BRANCH
No. 11 Jingu South Road
Huadong Town, Huadu District
Guangzhou, 510890 Guangdong, China
4. Youngyear Electronics Co., Ltd.
Jin Zhu Industries Area
Qing Xi Town, Dongguan City
Guandong
China

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date:

04.07.2019

Signature:



Dipl.-Ing. (FH) A. Klinker