

Certificate



Certificate no.

TA 50262994 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10041453 001

Client Reference: 1302034003/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Adapter)

License Fee - Units

Type Designation	: GSM40Ax (MW)	8
x stands for	: 05, 07, 09, 12, 15, 18, 24 or 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.0-0.5A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature:	45°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

9

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

12/08/2013



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50262994 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10041453 001

Client Reference: 1302034003/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Addition

Type Designation	: GSM60Ax (MW)	1
x stands for	: 05, 07, 09, 12, 15, 18, 24 or 48	1
Rated Current	: 1.4-0.7A	
max. Ambient Temperature:	40°C	

This certificate is valid for maximum 5 years.

2

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

12/08/2013



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50262994 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-SCL 10041453 001

Client Reference: 1302034003/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
12/08/2013

Certificate



Certificate no.

TA 50262994 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-SCL 10041453 001

Client Reference: 1302034003/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature




Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

12/08/2013

Certificate



Certificate no.

TA 50262994 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10041453 002

Client Reference: 05082014_A11

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)
22/08/2014

Certificate



Certificate no.

TA 50262994 06

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10041453 003

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

27/10/2015

Certificate



Certificate no.

TA 50262994 07

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10041453 004

Client Reference: 161001403

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Change


License Holder : see above

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/11/2016

Certificate



Certificate no.

TA 50262994 08

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1- 10041453 004

Client Reference: 161001403

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Change

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/11/2016

Certificate



Certificate no.

TA 50262994 09

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town,
Huadu District, Guangzhou 510890
P.R. China

Test report no.: ZTW1- 10041453 004

Client Reference: 161001403

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/11/2016

Certificate



Certificate no.

TA 50262994 10

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10041453 004

Client Reference: 161001403

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Adapter)

License Fee - Units

as page 01

Addition

Type Designation	: GSM40A13.5-KOD (MW)	1
Output Voltage	: DC 13.5V	
Output Current	: 2.96A	
max. Power Output	: 40W	
max. Ambient Temperature	: 45°C	
max. Operational Altitude	: 3000m	

This certificate is valid for maximum 5 years.

1

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/11/2016

Certificate



Certificate no.

TA 50262994 11

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI- 10041453 006

Client Reference: ES-18042017-TUV/GSM4

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (AC/DC Medical Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above
Product : see above

Addition

max. Operational Altitude: 3000m

Remark: 1. The risk management requirements of the standard
were not addressed;
2. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

25/01/2018