

Certificate



Certificate no.

TA 50457911 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuguan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50317323 001

Client Reference: 238012421-T-mark

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: MPM-x-yz (MW)	8
x stands for	: 45, 65, 90	1
y stands for	: 5, 12, 15, 24, 48	1
z stands for	: ST or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1) 1.2A (x= 45), 2) 1.5A (x= 65), 3) 1.9A (x= 90)	

Outputs : see constructional dataform
max. Ambient Temperature : 50°C (x= 45, 65), 45°C (x= 90)
max. Operational Altitude: 4000m

Protection Class : II

Remark: 1) The risk management requirements of the standard
were not addressed.

2) This certificate is valid for maximum 5 years.

11

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

20/01/2020

Anlage Fertigungsstättenliste
/Attachment List of Factories



TA 50457911 0001

1 MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

2 SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road, Dong-qiao,
Pan-yang Ind. Park, Huang-dai Town,
Xiang-cheng District Suzhou,
Jiangsu 215152
P. R. China

3 MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town, Huadu District,
Guangzhou 510890
P. R. China

Dieser Anhang ersetzt den Vorgänger vom/
This annex replaces the previous annex dated

Datum / Date 20.01.2020

Zertifizierungsstelle
/ Certification Body

Dipl.-Ing. W. Hsu

