

# Verification of Compliance

Product Name : **Medical Type Switching Power Supply**  
Brand Name : **MEAN WELL**  
Model Number : **MSP-450-x(x=3.3, 5, 7.5, 12, 15, 24, 36, 48)**  
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**  
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,  
Taiwan (R.O.C.)**  
Report Number : **S3C-M020-1303-108**  
Issue Date : **October 2, 2018**  
Applicable Standards :

**Emission :**

**EN 55011 : 2016+A1 : 2017  
Group 1, Class B  
EN 61000-3-2 : 2014  
EN 61000-3-3 : 2013**

**Immunity :**

**EN 60601-1-2 : 2015 Home Healthcare Environment  
EN 61000-4-2 : 2009  
EN 61000-4-3 : 2006+A1 : 2008+A2 : 2010  
EN 61000-4-4 : 2012  
EN 61000-4-5 : 2014  
EN 61000-4-6 : 2014  
EN 61000-4-8 : 2010  
EN 61000-4-11 : 2004**

Based on the EMC Directive 2014/30/EU, Medical Device Directive 93/42/EEC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905  
FCC CAB Code TW1104, TW0905  
NVLAP Lab Code 200575-0  
IC Code 4699A  
VCCI Accep. No. R-11527, C-11609, T-11441, G-10010,  
C-20010, T-20009, G-614



**Central Research Technology Co.**  
EMC Test Laboratory  
11, Lane41, Fushuen St., Jungshan Chiu,  
Taipei, Taiwan, 104, R.O.C.  
Tel : 886-2-25984568  
Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: October 2, 2018