

# Certificate



Certificate no.

TA 50433767 01

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd.,  
Wugu District, New Taipei City  
24891  
Taiwan

**Manufacturing Plant:**

Refer to latest revision  
of the annex list of factories

**Test report no.:** ZTW1- 50232419 001

**Client Reference:** 238005088

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Switching Mode Power Supply)

**License Fee - Units**

Type Designation	: RPS-120S-x (MW)	10
x stands for	: 12, 15, 24, 27, 48	1
Rated Voltage	: 1) AC 100-120V, 50/60Hz 2) AC 200-240V, 50/60Hz	
Rated Current	: 1) 2.3A 2) 1.1A	
Outputs	: see constructional dataform	
max. Ambient Temperature	: 50°C	
max. Operational Altitude	: 4000m	
Protection Class	: Not classified	

Continued on page 2

11

Appendix: 1

**Licensed Test mark:**



IEC 60601-1  
EN 60601-1

www.tuv.com  
ID 1419067964

**TÜV Rheinland Taiwan Ltd**

Signature

Dipl.-Ing. W. Hsu



**Date of Issue**

(day/mo/yr)

12/04/2019

Anlage Fertigungsstättenliste  
/Attachment List of Factories



TA 50433767 0001

1 MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd.,  
Wugu District, New Taipei City  
24891  
Taiwan

2 SuZhou MEAN WELL Technology  
Co., Ltd.  
No. 77, Jian-min Road, Dong-qiao, Pan-  
yang Ind. Park, Huang-dai Town, Xiang-  
cheng District Suzhou, Jiangsu 215152  
P. R. China

3 MEAN WELL (GUANGZHOU) ELECTRONICS  
CO., LTD HUADU BRANCH  
No. 11 Jingu South Road,  
Huadong Town, Huadu District,  
Guangzhou 510890  
P. R. China

TÜVRheinland

Dieser Anhang ersetzt den Vorgänger vom/  
This annex replaces the previous annex dated

Datum / Date 12.04.2019

Zertifizierungsstelle  
/ Certification Body

Dipl.-Ing. W. Hsu



# Certificate



Certificate no.

TA 50433767 02

**License Holder:**

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd.,  
Wugu District, New Taipei City  
24891  
Taiwan

**Manufacturing Plant:**

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd.,  
Wugu District, New Taipei City  
24891  
Taiwan

**Test report no.:** ZTW1- 50232419 001

**Client Reference:** 238005088

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Switching Mode Power Supply)

**License Fee - Units**

as page 01, Continuation

- Remark:
- 1) The risk management and usability requirements of the standard were not addressed.
  - 2) The certified product was not classified. The actual classification shall be determined for the end product assembly.
  - 3) This certificate is valid for maximum 5 years.

Appendix: 1

**Licensed Test mark:**



IEC 60601-1  
EN 60601-1

www.tuv.com  
ID 1419067964

**TÜV Rheinland Taiwan Ltd**

Signature

Dipl.-Ing. W. Hsu



**Date of Issue**

(day/mo/yr)

12/04/2019